CAUTION: READ ENTIRE PACKAGE INSERT FOR IMPORTANT STORAGE AND USE INFORMATION. STORE IN A COOL DRY PLACE.

STORAGE

Store the Rapicide glutaraldehyde indicator test strips at room temperature between 43°F and 86°F (6°C and 30°C) in the original bottle with the cap tightly closed and in a dry place. Do not refrigerate or freeze.

IMPORTANT PRECAUTIONS - Keep the test strips in the original container. Remove only the required number of test strips before testing. After removing the test strip from the bottle immediately replace the cap making sure it is on tight. Do not freeze or expose to light, heat or moisture, and do not use test strips 90 days after opening the bottle. Do not expose to aldehydes, strong acids/alkalis and detergents. Do not use the test strips after the expiration date (opened or unopened). Do not touch the reagent pad with fingers or any other surface.

Rapicide®

GLUTARALDEHYDE INDICATOR TEST STRIPS

Potency Test Strips for Glutaraldehyde High Level Disinfectant Solution Monitoring

60 Strips

Cat. No. ML02-0120

DATE OPENED

DIRECTIONS FOR USE

Refer to the endoscope reprocessor's user manual for instructions on how to take a sample of the disinfectant use solution.

- Dip the strip into the solution for exactly (3) three seconds and immediately remove.
- Remove excess solution by touching short edge of test strip to paper towel. Do not shake or blot the test strip.
- Wait for exactly (75) seventy-five seconds and then compare the test strip color with the color chart on the bottle to determine the glutaraldehyde level.

Manufactured in the UK for:

Endoscope reprocessing the way it should be™ MEDIVATORS®

Minntech Corporation
14605 28th Avenue North
Minneapolis, Minnesota 55447 USA
Toll Free: 1-800-444-4729
www.minntech.com

LOT :

EXP:

67199-227/D